

**CERTIFICATE OF COMPLIANCE FOR RADIOACTIVE MATERIAL PACKAGES
D.O.T. SPECIFICATIONS 7A, TYPE A**

**Model #001-724 PET Vial Shipping System
Page 1 of 2**

1. Package Identification:

Common Name: Single or Multiple Dose Shipping System
Ground or Air Shipment
Catalog Name: **Shipping System, PET, Vial**
Model Number: **#001-724**

2. Preamble:

- a. This certificate is issued to satisfy section 178.350 including sections 171.2, 173.403, 173.410, 173.411, 173.412, 173.415, 173.431, 173.465, 173.466 and 178.608 of the Department of Transportation Hazardous Materials Regulations (49 CFR).
- b. The packaging and contents described in item 4 below were found to meet the safety standards set forth in, and tested, according to the more stringent recommendations for general performance testing from:
- IATA Dangerous Goods Regulations, 60th Edition Sections 5.0.4.3, 10.5, and 10.6.1 thru 10.6.3.5.
- 49 CFR as listed in item 2a above.
- c. This certificate does not release the consignor from compliance with any requirement of the regulations of the U.S. Department of Transportation or other applicable regulatory agencies, including the government of any country through or into which the package will be transported.

3. Descriptions of Packaging and Authorized contents, Other Conditions, and References:

- a. **Outer Container:** Model #001-783-A005
Medium Density Polyethylene:
Dimension: 11.75" x 11.75" x 12.5" with high density polyurethane foam cushioning and insert support
Weldment with two locking twist hasps and a hinged lid. Case has two handles. A safety seal can be put thru the hasp and the stand off for tamper proof evidence.
- b. **Shielded Insert:** Model #001-724-C200
Contour lead shield with open top and bottom and 3 pieces of ¼" lead in the bottom of the case. Contoured lead thickness varies from 1.232" to 1.699".
- c. **PET Unit Dose Pig:** Model #001-706
0.065" steel with one inch of lead in the walls and bottom of the pig
Top made of aluminum with 1.75" of lead shielding
- d. **Combined Lead Shielding:** Bottom = 1.75"
Top = 1.75"
Sides: Varies depending on location from 1.232" to 1.699"

**CERTIFICATE OF COMPLIANCE FOR RADIOACTIVE MATERIAL PACKAGES
D.O.T. SPECIFICATIONS 7A, TYPE A**

**Model #001-724 PET Vial Shipping System
Page 2 of 2**

- e. **Absorbent Material:** Model #001-771
One piece of 3" x 3" Quick Solid absorbent material in the bottom of each pig
- f. **Vial:** 30ml Abbott Lab Vial or 10ml vial with plastic vial adapter

Authorized Contents:

Normal Form Radioactive Materials in Type A quantities
Tested with F-18, 1.5 curies in the PET Vial Pig inside the case as Yellow II. Must comply with the A, Value Limits
From 49CFR 173.435.

Specifications and Restrictions:

Ground or Air Transportation

4. This certificate is issued on the basis of the test results and safety analysis of the package design.

5. Certification

This document is to certify that to the best of our knowledge, information and belief, the construction methods, the packaging design, and the materials of construction for this specification package comply with the standards described in items 2a, 2b, 2c herein. Any deviation from these conditions must be approved prior to use.

Approved:



Clyde Schlein
Vice President, Regulatory Affairs & Compliance
Biodex Medical Systems, Inc.

Date: Jan 4, 2013
Rev: Jan 22, 2019