

Products included in the Certificate No: 41313009-01
 Issued to: **Biodex Medical Systems, Inc.**
 20 Ramsey Road
 11967 Shirley, New York
 USA

Product category	Type/Model designation	Class	Measuring	GMDN code <small>(not mandatory)</small>	Date added
Physical Therapy and Rehabilitation Equipment	System 4	Ila	No	-	*
Medical Spectrometers	Atomlab 960 187-600, 187-601	Ila	No	-	Jan 25, 2013

* Product added before January 25, 2013.

Signed Date: 19 July 2018
 Valid Date: 21 July 2018

Intertek Semko AB
 Notified Body MDD



Pontus Gedda
 Certification Authority MDD

This product list is only valid together with the referenced, valid EC certificate.

The GMDN codes are assigned by the manufacturer and are only provided for convenience.

Intertek Semko AB is a Notified Body according to the Directive 93/42/EEC on medical devices, with identification number 0413.

Certificate No: 41313009-01
Date: 19 July 2018
Handled by: Caroline Åman
E-mail: medtechsweden@intertek.com

Biodex Medical Systems, Inc

Attn: Clyde Schlein
20 Ramsey Road
11967-0702 Shirley, New York
USA

Purpose	Assessment to issue a new certificate due to five year extension according to the national legislation for medical devices LVFS 2003:11 (Medical Device Directive 93/42/EEC), Annex II.
Activity	Certification audit was performed 26 June 2017 in Shirley, New York by Richard Auringer and Luis Lopes The technical file was reviewed by Lian Zhang at Intertek's office.
Scope of assessment	- Physical therapy and rehabilitation equipment - Medical spectrometers, Class IIa
Result	2 minor non conformities were noted during the audit. Presented corrective action plans have been examined and approved by us.
Certificate Valid from	21 July 2018
Conclusions/Decisions	Referring to the above a Certificate of Conformance with the national legislation for medical devices LVFS 2003:11 (Medical Device Directive 93/42/EEC), Annex II will be issued. The Certificate is valid for products specified in the "MDD – Product List".
Follow-up assessments	Follow-up assessments are going to be performed once a year.
Appeals	Any appeal against this decision will be processed by an appeals panel as Intertek. The appeal shall be submitted to Intertek Semko AB, PO-Box 1103, SE-164 22 Kista, Sweden.
Others	Any complaints, from customers and others, and corrective actions concerning your certified quality system shall be documented and retained. Upon request Intertek Semko has the right to review this documentation.

Intertek Semko AB
Notified Body MDD



Pontus Gedda
Certification Authority MDD